Greetings from Lumea. We are glad you are interested in the BxChip.

First, check out this video:

Next, order a **Demo Kit** from orders@lumeadigital.com (mailto:orders@lumeadigital.com)

We can schedule a call with you to explain the kit and answer any other questions about the BxChip.

Once you have some experience playing with the BxChip, we recommended ordering a **Validation Kit** and following the instructions in the BxChip Validation Process (https://resources.lumeadigital.com/resource/bxchip-validation-process/) .

Key Points

- Validate the BxChip in your tissue processor before using on patient samples. Reference the validation process (https://resources.lumeadigital.com/resource/bxchip-validation-process/) for more information.
- Do not leave the BxChip out of formalin/saline for more than 15 minutes prior to processing. It will dry out!
- Embed the BxChip with the grooves facing **down** in the embedding mold.
- Do not get the BxChip block wet. Use solid ice, freezer, etc. to cool the block prior to microtomy.